



ESWW-15 HSDC TDM

EPT recalibration achievements

by

S. Borisov, S. Benck and M. Cyamukungu
(UCLouvain/ELI/CSR)


7 November 2018
Leuven




SEP-2017/09 case study

- S1 overcounting problem was discovered 1 year ago;
- Strong filtering in order to suppress improper measurements is developed;
- Comparison to the GOES (effective energies by Sandberg et al.) shows good agreement.

11.2 MeV H PROBA-V/EPT (red) and GOES-13 (black) flux



45.0 MeV H PROBA-V/EPT (red) and GOES-13 (black) flux



109 MeV H PROBA-V/EPT (red) and GOES-13 (black) flux

